


Application Note: 57-AM

## Isocratic Separation of Amphenicols on HALO Phenyl-Hexyl Phase




### PEAK IDENTITIES:

1. Thiamphenicol
2. Chloramphenicol


### TEST CONDITIONS:

Column: 4.6 x 50 mm, HALO Phenyl-Hexyl  
 Part Number: 92814-406  
 Mobile Phase: 55/45-A/B  
 A= 0.025 M Ammonium acetate buffer, pH=5.8  
 B=Acetonitrile  
 Flow Rate: 1.0 mL/min.  
 Pressure: 94 Bar  
 Temperature: 35 °C  
 Detection: UV 240/280 nm, VWD  
 Injection Volume: 0.3 µL  
 Sample Solvent: Acetonitrile  
 Response Time: 0.02 sec.  
 Flow Cell: 2.5 µL semi-micro  
 LC System: Shimadzu Prominence UFLC XR  
 Extra column volume: ~14 µL

### STRUCTURES:



Thiamphenicol



Chloramphenicol

This separation shows a rapid HPLC method for the analysis of amphenicols on HALO Phenyl-Hexyl stationary phase. To improve the sensitivity of detection the first peak was monitored @ 240 nm and the second @ 280 nm.